

```
SmartSelect = "yes";
Coil 1 (exclude) = "None";
Uniformity = "CLEAR";
FOV          FH (mm) = 256;
              AP (mm) = 256;
              RL (mm) = 211.200012;
ACQ voxel size  FH (mm) = 1;
                  AP (mm) = 1;
                  RL (mm) = 1.20000005;
Recon voxel size FH (mm) = 1;
                  AP (mm) = 1;
                  RL (mm) = 1.20000005;
Fold-over suppression = "no";
Slice oversampling = "user defined";
    oversample factor = 1;
RF select. FOS = "no";
ENCASE enable = "no";
Reconstruction matrix = 256;
SENSE = "yes";
    P reduction (AP) = 2;
    S reduction (RL) = 1;
fMRI = "no";
Z Shim = "no";
k-t BLAST = "no";
Stacks = 1;
```

```
slices = 176;
slice orientation = "sagittal";
fold-over direction = "AP";
fat shift direction = "F";
Stack Offc. AP (P=+mm) = 0;
        RL (L=+mm) = 0;
        FH (H=+mm) = 0;
Ang. AP (deg) = 0;
        RL (deg) = -0;
        FH (deg) = -0;
Free rotatable = "no";
Multi-chunk = "no";
Large table movement = "no";
PlanAlign = "no";
REST slabs = 0;
Interactive positioning = "no";
External control = "no";
Patient position = "head first";
Patient body position = "head first";
Patient orientation = "supine";
Patient body orientation = "supine";
Scan type = "Imaging";
Scan mode = "3D";
        technique = "FFE";
Contrast enhancement = "no";
```

```
Acquisition mode = "cartesian";
Fast Imaging mode = "TFE";
    3D non-selective = "no";
    shot mode = "multishot";
TFE factor = 176;
    3D free factor = "no";
    startup echoes = "default";
    shot interval = "user defined";
        (ms) = 2530;
    profile order = "linear";
    turbo direction = "Z";
Echoes = 4;
    partial echo = "no";
    shifted echo = "no";
TE first = "shortest";
    echospacing = "user defined";
        (ms) = 1.86000001;
    flyback = "no";
Flip angle (deg) = 7;
TR = "shortest";
Halfscan = "no";
Water-fat shift = "user defined";
    (pixels) = 0.660000026;
Shim = "auto";
mDIXON = "no";
```

Fat suppression = "no";
Water suppression = "no";
TFE prepulse = "invert";
 slice selection = "no";
 delay = "user defined";
 (ms) = 1300;
 PSIR = "no";
MTC = "no";
T2prep = "no";
Research prepulse = "no";
Diffusion mode = "no";
Multi-transmit = "no";
SAR mode = "high";
B1 mode = "default";
SAR allow first level = "yes";
PNS mode = "high";
Gradient mode = "maximum";
SoftTone mode = "no";
Cardiac synchronization = "no";
Heart rate > 250 bpm = "no";
Respiratory compensation = "no";
Navigator respiratory comp = "no";
Flow compensation = "no";
fMRI echo stabilisation = "no";
Motion smoothing = "no";

NSA = 1;
Angio / Contrast enh. = "no";
Quantitative flow = "no";
CENTRA = "no";
Manual start = "no";
Dynamic study = "no";
Arterial Spin labeling = "no";
Preparation phases = "auto";
Interactive F0 = "no";
B0 field map = "no";
B1 field map = "no";
MIP/MPR = "no";
SWIp = "no";
Images = "M", (3) "no";
Autoview image = "P";
Calculated images = (4) "no";
Reference tissue = "Grey matter";
Recon compression = "No";
Preset window contrast = "soft";
Reconstruction mode = "immediate";
Save raw data = "no";
Hardcopy protocol = "no";
Image filter = "system default";
Uniformity correction = "no";
Geometry correction = "default";

Elliptical k-space shutter = "default";
IF_info_seperator = 1634755923;
Total scan duration = "05:26.0";
Rel. SNR = 1;
Act. TR/TE1/delta TE (ms) = "9.2 / 1.65 / 1.9";
ACQ matrix M x P = "256 x 256";
ACQ voxel MPS (mm) = "1.00 / 1.00 / 1.20";
REC voxel MPS (mm) = "1.00 / 1.00 / 1.20";
Scan percentage (%) = 100;
Act. slice gap (mm) = 0;
TFE shots = 128;
TFE dur. shot / acq (ms) = "2123.6 / 1628.1";
Min. TI delay = 854.735291;
Act. WFS (pix) / BW (Hz) = "0.667 / 651.0";
Min. WFS (pix) / Max. BW (Hz) = "0.329 / 1319.6";
Head SAR = "< 6 %";
Whole body SAR / level = "0.0 W/kg / normal";
SED = " 0.0 kJ/kg";
Max B1+rms = "0.58 uT";
PNS / level = "84 % / 1st level";
dB/dt = "111.0 T/s";
Sound Pressure Level (dB) = 21.109024;