

Nucleus = "H1";
SmartSelect = "yes";
Coil 1 (exclude) = "None";
Uniformity = "CLEAR";
FOV RL (mm) = 220;
AP (mm) = 220;
FH (mm) = 136.800003;
Voxel size RL (mm) = 3.4375;
AP (mm) = 3.4375;
Slice thickness (mm) = 3.79999995;
Recon voxel size (mm) = 3.4375;
Fold-over suppression = "no";
Reconstruction matrix = 64;
SENSE = "yes";
P reduction (AP) = 2;
k-t BLAST = "no";
Stacks = 1;
type = "parallel";
slices = 36;
slice gap = "user defined";
gap (mm) = 0;
slice orientation = "transverse";
fold-over direction = "AP";
fat shift direction = "P";
Stack Offc. AP (P=+mm) = -11.0033407;
RL (L=+mm) = 0.848175883;
FH (H=+mm) = 21.4465027;
Ang. AP (deg) = 4.01817942;
RL (deg) = -8.74200058;
FH (deg) = 1.47300553;
Free rotatable = "no";
Minimum number of packages = 1;
Slice scan order = "ascend";
Large table movement = "no";
PlanAlign = "no";
REST slabs = 0;
Catheter tracking = "no";
Interactive positioning = "no";
Allow table movement = "no";
Patient position = "head first";
orientation = "supine";
Scan type = "Imaging";
Scan mode = "MS";
technique = "FFE";
Contrast enhancement = "no";
Acquisition mode = "cartesian";
Fast Imaging mode = "EPI";
shot mode = "single-shot";
Echoes = 1;
partial echo = "no";
shifted echo = "no";
TE = "user defined";
(ms) = 21;
Flip angle (deg) = 90;

TR = "user defined";
(ms) = 1500;
Halfscan = "no";
Water-fat shift = "user defined";
(pixels) = 8;
Shim = "auto";
mDIXON = "no";
Fat suppression = "SPIR";
strength = "strong";
frequency offset = "default";
Water suppression = "no";
MTC = "no";
Research prepulse = "no";
Diffusion mode = "no";
+Edit duty cycle = "yes";
Elastography mode = "no";
SAR mode = "high";
B1 mode = "default";
SAR Patient data = "auto";
PNS mode = "high";
Gradient mode = "maximum";
SoftTone mode = "no";
Cardiac synchronization = "no";
Heart rate > 250 bpm = "no";
Respiratory compensation = "no";
Navigator respiratory comp = "no";
Flow compensation = "no";
Temporal slice spacing = "default";
fMRI echo stabilisation = "no";
NSA = 1;
Angio / Contrast enh. = "no";
Quantitative flow = "no";
Manual start = "no";
Dynamic study = "individual";
dyn scans = 281;
recon multiplier = 1;
dyn scan times = "shortest";
fov time mode = "default";
dummy scans = 6;
immediate subtraction = "no";
fast next scan = "no";
+Physiology marker = "no";
synch. ext. device = "no";
dyn stabilization = "regular";
prospect. motion corr. = "no";
Keyhole = "no";
Arterial Spin labeling = "no";
Preparation phases = "full";
Interactive F0 = "no";
SmartPlan survey = "no";
B0 field map = "no";
B1 field map = "no";
MIP/MPR = "no";
SWIp = "no";

Images = "M", (3) "no";
Autoview image = "M";
Calculated images = (4) "no";
Reference tissue = "Grey matter";
EPI 2D phase correction = "no";
Recon compression = "No";
Preset window contrast = "soft";
Reconstruction mode = "real time";
reuse memory = "no";
Save raw data = "no";
Hardcopy protocol = "no";
Image filter = "system default";
Geometry correction = "default";
IF_info_seperator = 0;
Total scan duration = "07:13.4";
Rel. SNR = 1;
Act. TR/TE (ms) = "1500 / 21";
Dyn. scan time = "00:01.5";
Time to k0 = "0.744";
ACQ matrix M x P = "64 x 62";
ACQ voxel MPS (mm) = "3.44 / 3.55 / 3.80";
REC voxel MPS (mm) = "3.44 / 3.44 / 3.80";
Scan percentage (%) = 96.875;
Packages = 1;
Min. slice gap (mm) = -0;
EPI factor = 31;
Act. WFS (pix) / BW (Hz) = "8.000 / 54.3";
BW in EPI freq. dir. (Hz) = "2418.0";
Min. WFS (pix) / Max. BW (Hz) = "5.338 / 81.3";
Min. TR/TE (ms) = "1482 / 7.4";
SAR / local torso = "< 100 %";
Whole body / level = "< 1.3 W/kg / normal";
SED = "< 0.6 kJ/kg";
B1+rms = "1.61 uT";
Max B1+rms = "1.62 uT";
PNS / level = "92 % / 1st level";
dB/dt = "119.3 T/s";
Sound Pressure Level (dB) = 21.068655;